



Calibration  
Certificate No. 1750.01

Calibration complies with ISO/IEC  
17025, ANSI/NCSL Z540-1, and 9001



Cert. No.: 4227-7602387

**Traceable® Certificate of Calibration for Refrigerator/Freezer Thermometer**

Manufactured for and distributed by: Fisher Scientific, 300 Industry Drive, Pittsburgh, PA 15275-1001

**Instrument Identification:**

Model Numbers: 06-664-23, 11705853 S/N: 160380145 Manufacturer: Control Company

**Standards/Equipment:**

Description	Serial Number	Due Date	NIST Traceable Reference
Temperature Calibration Bath TC-179	A45240		
Readout, Digital Thermometer	B5C344	12/30/16	B5C30097
Thermistor Probe	5357	12/14/16	B5C07066
Temperature Calibration Bath TC-256	B01375		
Thermistor Probe	5356	12/15/16	B5C07065
Readout, Digital Thermometer	B5C344	12/30/16	B5C30097

**Certificate Information:**

Technician: 68 Procedure: CAL-03 Cal Date: 4/20/16 Due Date: 4/20/18  
Test Conditions: 24.3°C 47.0 %RH 1015 mBar

**Calibration Data: (New Instrument)**

Unit(s)	Nominal	As Found	In Tol	Nominal	As Left	In Tol	Min	Max	±U	TUR
°C Probe		N.A.		0.000	0.1	Y	-0.5	0.5	0.059	>4:1
°C Probe		N.A.		50.000	50.1	Y	49.5	50.5	0.059	>4:1

**This Instrument was calibrated using Instruments Traceable to National Institute of Standards and Technology.**

A Test Uncertainty Ratio of at least 4:1 is maintained unless otherwise stated and is calculated using the expanded measurement uncertainty. Uncertainty evaluation includes the instrument under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM). The uncertainty represents an expanded uncertainty using a coverage factor k=2 to approximate a 95% confidence level. In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The results contained herein relate only to the item calibrated. This certificate shall not be reproduced except in full, without written approval of Control Company.

Nominal=Standard's Reading; As Left=Instrument's Reading; In Tol=In Tolerance; Min/Max=Acceptance Range; ±U=Expanded Measurement Uncertainty; TUR=Test Uncertainty Ratio; Accuracy=±(Max-Min)/2; Min = As Left Nominal(Rounded) - Tolerance; Max = As Left Nominal(Rounded) + Tolerance; Date=MM/DD/YY

This certificate indicates calibration for external sensor only.

*Nicol Rodriguez*  
Nicol Rodriguez, Quality Manager

*Aaron Judice*  
Aaron Judice, Technical Manager

**Maintaining Accuracy:**

In our opinion once calibrated your Refrigerator/Freezer Thermometer should maintain its accuracy. There is no exact way to determine how long calibration will be maintained. Refrigerator/Freezer Thermometers change little, if any at all, but can be affected by aging, temperature, shock, and contamination.

**Recalibration:**

For factory calibration and re-certification traceable to National Institute of Standards and Technology contact Control Company.

**CONTROL COMPANY 4455 Rex Road Friendswood, TX 77546 USA**  
**Phone 281 482-1714 Fax 281 482-9448 service@control3.com www.control3.com**

Control Company is an ISO/IEC 17025:2005 Calibration Laboratory Accredited by (A2LA) American Association for Laboratory Accreditation, Certificate No. 1750.01.  
Control Company is ISO 9001:2008 Quality Certified by DNV GL, Certificate No. CERT-01805-2006-AQ-HOU-RvA.  
International Laboratory Accreditation Cooperation (ILAC) - Multilateral Recognition Arrangement (MRA).