



Calibration complies with ISO/IEC
17025, ANSI/NCSL Z540-1, and 9001



Cert. No.: 4227-9907136

Traceable® Certificate of Calibration for Refrigerator/Freezer Thermometer

Manufactured for and distributed by : Fisher Scientific "300 Industry Drive,,Pittsburgh,PA,15275-1001"

Instrument Identification:

Model: 06-664-23,11705853

S/N: 181694020

Manufacturer: Control Company

Standards/Equipment:

Description	Serial Number	Due Date	NIST Traceable Reference
Temperature Calibration Bath	B3A444		
Digital Thermometer	140156092	17 Jul 2019	4000-9654118
Temperature Calibration Bath	B5C478		
Digital Thermometer	170641111	19 Sep 2019	6412-9809961

Certificate Information:

Technician: 104

Procedure: CAL-03

Cal Date: 25 Oct 2018

Cal Due Date: 25 Oct 2020

Test Conditions: 54.04%RH 23.57°C 1009mBar


Calibration Data: (New Instrument)


Unit(s)	Nominal	As Found	In Tol	Nominal	As Left	In Tol	Min	Max	±U	TUR
°C Probe	N.A.	N.A.		0.00	0.0	Y	-0.5	0.5	0.058	>4:1
°C Probe	N.A.	N.A.		50.00	49.9	Y	49.5	50.5	0.058	>4:1

This certificate indicates Traceability to standards provided by (NIST) National Institute of Standards and Technology and/or a National Standards Laboratory.

A Test Uncertainty Ratio of at least 4:1 is maintained unless otherwise stated and is calculated using the expanded measurement uncertainty. Uncertainty evaluation includes the instrument under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement : (GUM). The uncertainty represents an expanded uncertainty using a coverage factor k=2 to approximate a 95% confidence level. In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The results contained herein relate only to the item calibrated. This certificate shall not be reproduced except in full, without written approval of Control Company.

Nominal=Standard's Reading; As Left=Instrument's Reading; In Tol=In Tolerance; Min/Max=Acceptance Range; ± U=Expanded Measurement Uncertainty; TUR=Test Uncertainty Ratio; Accuracy=±(Max-Min)/2; Min=As Left Nominal(Rounded) - Tolerance; Max= As Left Nominal(Rounded) + Tolerance;


Nicol Rodriguez, Quality Manager


Aaron Judice, Technical Manager

Note :

Maintaining Accuracy:

In our opinion once calibrated your Refrigerator/Freezer Thermometer should maintain its accuracy. There is no exact way to determine how long calibration will be maintained. Refrigerator/Freezer Thermometer change little, if any at all, but can be affected by aging, temperature, shock, and contamination.

Recalibration:

For factory calibration and re-certification traceable to National Institute of Standards and Technology contact Control Company.

CONTROL COMPANY 12554 Galveston RD Suite B230 Webster TX USA 77598
Phone 281 482-1714 Fax 281 482-9448 sales@control3.com www.control3.com

Control Company is an ISO/IEC 17025:2005 Calibration Laboratory Accredited by (A2LA) American Association for Laboratory Accreditation, Certificate No. 1750.01.
Control Company is ISO 9001:2008 Quality Certified by DNV GL, Certificate No. CERT-01805-2006-AQ-HOU-RvA.
International Laboratory Accreditation Cooperation (ILAC) - Multilateral Recognition Arrangement (MRA).