

SENSITECH INC.
CERTIFICATE OF CALIBRATION



Sensitech Inc. • 800 Cummings Center • Suite 258X • Beverly, MA 01915-6197 • USA: 1-800-843-8367 • Int'l: +1-978-927-7033

PRODUCT TempTale®
STATEMENT OF CALIBRATION Sensitech Inc. certifies that the device(s) on the Device List have been thoroughly tested and calibrated in accordance with Sensitech procedures and meet or exceed the accuracy specifications*.
Calibration expires one year from the date of sale.
*Specification details for Sensitech products are provided in the published datasheets.
REFERENCE EQUIPMENT** Fluke Hart Scientific thermometer model 1502A /Burns Engineering PRT model 3925 or equivalent
Temperature Accuracy: $\pm 0.05^{\circ}\text{C}$; Calibration Range: -195°C to $+232^{\circ}\text{C}$

**Reference equipment calibration is accredited to ISO/IEC 17025:2005 and is traceable through the National Institute of Standards and Technology (NIST), or other National Standards Institutes, to the International System of Units (SI). Calibration certificates for all reference equipment are retained by Sensitech Quality Assurance at our headquarters in Beverly, Massachusetts, USA.

ATTESTED BY Director of Quality Assurance

Scott McDowell

5 Dec 2015

Name

Signature

Date

SHIP TO Carrier Transicold Hong Kong Ltd
7/F Chung Shun Knitting Centre
1-3 Wing Yip Street
Kwai Chung New Territories Hong Kong
SOLD TO Carrier Transicold Hong Kong Ltd
7/F Chung Shun Knitting Centre
1-3 Wing Yip Street
Kwai Chung New Territories Hong Kong

ORDER NUMBER JY5101868

ORDER LINE NUMBER 19

ORDER RELEASE NUMBER 1

PACK ID N/A

RELEASED FROM Hong Kong Facility

RELEASED BY

John Wong

23-Jul-2021

Name

Signature

Date

DATE OF SALE 2021-07-27

DEVICE LIST



SHIP TO Carrier Transicold Hong Kong Ltd 7/F Chung Shun Knitting Centre 1-3 Wing Yip Street Kwai Chung New Territories Hong Kong		SOLD TO Carrier Transicold Hong Kong Ltd 7/F Chung Shun Knitting Centre 1-3 Wing Yip Street Kwai Chung New Territories Hong Kong	ORDER NUMBER JY5101868
			ORDER LINE NUMBER 19
			ORDER RELEASE NUMBER 1
			PACK ID N/A
Pack#: BOX ID:	N/A 21148H0202	ORDER LINE: 19	JY5101868 RELEASE 1
SERIAL NUMBER	MODEL NUMBER	DESCRIPTION	CALIBRATION DATE
JFJ4N04ZJ0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N054W0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N056X0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N057H0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05J10	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05J50	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05J60	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05JB0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05JH0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05JK0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05KZ0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05T10	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05VB0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05VC0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05VD0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05VQ0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05VV0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05VX0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05W10	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05W20	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021

室温 →

冷 I →

冷 II →

DEVICE LIST



SHIP TO Carrier Transicold Hong Kong Ltd 7/F Chung Shun Knitting Centre 1-3 Wing Yip Street Kwai Chung New Territories Hong Kong		SOLD TO Carrier Transicold Hong Kong Ltd 7/F Chung Shun Knitting Centre 1-3 Wing Yip Street Kwai Chung New Territories Hong Kong	ORDER NUMBER JY5101868
			ORDER LINE NUMBER 19
			ORDER RELEASE NUMBER 1
			PACK ID N/A
Pack#: N/A BOX ID: 21148H0203		ORDER JY5101868 LINE: 19 RELEASE 1	
SERIAL NUMBER	MODEL NUMBER	DESCRIPTION	CALIBRATION DATE
JFJ4N04X00	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N04ZX0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05AK0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05AR0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05AW0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05CG0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05CH0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05E30	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05E70	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05EB0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05EC0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05ED0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05EF0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05K70	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05VA0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05VF0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05VP0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05W60	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05WD0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021
JFJ4N05WG0	TUA00-01	TEMPTALE ULTRA 16K AMB SU	23-Jul-2021

冷Ⅲ →
冷Ⅳ →
冷Ⅴ →

Line/Rel Total: 40